

TD-4129

BLOOD GLUCOSE MONITORING SYSTEM

5 Sec, 1.0 μ L

Display Screen



FEATURES

- Display both glucose level and HCT level
- Strip eject function
- Before / After meal (AC/PC) presentable
- 7, 14, 21, 28, 60, 90 day average function
- RS232 data transmission



SPECIFICATIONS

Enzyme Type	GDH-FAD
Sample Size	1.0 μ L
Reaction Time	5 seconds
Measurement Range	10~600mg/dL (0.55~33.3mmol/L)
Hematocrit Range	0% - 70%
Precision	SD < 5 mg/dL (0.278 mmol/L) at < 100mg/dL (5.55 mmol/L); CV < 5% at 100mg/dL (5.55 mmol/L)
Accuracy	\pm 15mg/dL if < 100mg/dL : \pm 15% if \geq 100mg/dL
Ketone Warning	Yes
Communication	RS232-4 Poles
Power Source	CR2032 * 1
Memory Capacity	420 sets
Day Average	7, 14, 21, 28, 60, 90 days
Daily Alarm	4 daily alarms
Dimension	98 (L) x 54 (W) x 15 (H) mm
Weight	48 g (without battery)
Operating Condition	10°C(50°F) ~ +40°C(104°F), below 85% R.H.
Storage Condition	-20°C(-4°F) ~ +60°C(140°F) (Meter); 2°C(35.6°F)~32°C(89.6°F) (Strip)
Units Per Carton	30
Carton Size	340 (L) x 290 (W) x 383 (H) mm
Cross Weight Per Carton	11.1 kg



www.taidoc.com

TaiDoc Technology Corp.

B1-7F., No.127, Wugong 2nd Rd., Wugu Dist.,
New Taipei City 24888, Taiwan
Tel : +886-2-6625-8188 Fax : +886-2-6625-0288